

TECHNICAL DATA SHEET

Article no. : 8536.0500

KANAMYCIN ESCULIN AZIDE AGAR (KAA AGAR)

SPECIFICATION

Solid medium for confirmative detection and isolation of Lancefield's group D streptococci in food samples, according to Mossel *et al.*

FORMULA * IN G/L

| | |
|------------------------------|-------|
| Tryptone..... | 20.00 |
| Yeast extract..... | 5.00 |
| Sodium chloride..... | 5.00 |
| Disodium citrate..... | 1.00 |
| Esculin..... | 1.00 |
| Ferric-ammonium citrate..... | 0.50 |
| Sodium azide..... | 0.15 |
| Kanamycin sulfate..... | 0.02 |
| Agar..... | 15.00 |

Final pH 7,0 ±0,2 at 25 °C

*Adjusted and /or supplemented as required to meet performance criteria

DIRECTIONS

Suspend 48 g of powder in 1 L of distilled water and let it soak. Heat to boiling point and distribute into suitable containers. Sterilize in the autoclave at 121°C for 15 minutes.

DESCRIPTION

KAA confirmative Agar is a medium that several organisations and institutes recommend for detecting, enumerating and isolating Lancefield's group D streptococci in samples of food and beverages e.g.: bottled water, fresh/refrigerated/frozen/minced meat, fish, molluscs, soft drinks, pastries and spices. Kanamycin and sodium azide are the selective inhibitory compounds.

TECHNIQUE

From samples considered positive, aliquots of 0,1 mL are inoculated onto the surface of the plates of KAA, spreading with a Drigalsky loop. Incubate the plates, in an inverted position, at 36°C for 24-48 hours. Colonies that appear surrounded by a black halo are considered as group D streptococci, and are isolated to confirm them biochemically and morphologically with the following tests: microscopical examination; catalase assay (that should be negative) in an azide-less medium; growth at 45°C and resistance to a high saline concentration (6,5% of NaCl in BHI Broth).

Finally, they have to grow in Bile Esculin Agar with an appearance similar to the colonies on the KAA Confirmative Agar. Nonetheless, there are some exceptions to this rule, i.e. *Streptococcus equinus* and *S. bovis* do not grow in the hypersaline broth, and therefore, definitive identification has to be performed by serological methods.

This methodology does not allow the enumeration of bacteria from the original sample, and as this is a necessary, the Most Probable Number (MPN) technique is recommended with KAA Presumptive Broth, using double strength broth if necessary.



Th. Geyer GmbH & Co. KG
Domierstr. 4 – 8
D-71272 Renningen
Tel.: +49 7159 1637-0
Fax: +49 7159 1637-710
rennigen@thgeyer.de
www.thgeyer.de

BW-Bank (Swift/BIC SOLADEST800)
IBAN DE85800501010002036302
Postbank Stuttgart (Swift/BIC PBNKDEFFXXX)
IBAN DE32800100700000020708
Deutsche Bank (Swift/BIC DEUTDESSXXX)
IBAN DE06800700700125518100

St.-Nr. 70093/40018 / USt-IdNr. DE147510304
Amtsgericht Stuttgart / HRA-Nr. 254140
Persönlich haftende Gesellschafterin:
Geyer Beteiligungsgesellschaft mbH
Amtsgericht Stuttgart / HRB-Nr. 252035
Geschäftsführer: Lutz-Alexander Geyer /
Oliver-Alexander Geyer / André Meise / Ralf Streicher

QUALITY CONTROL

Incubation Temperature: 36°C ±2.0

Incubation Time: 44 ± 4h

Inoculum: 10³-10⁴ CFU (Productivity test qualitative)/ 10⁴-10⁶ CFU (Selectivity) according to ISO 11133:2014/Amd 1:2018.

| Microorganism | Growth | Remarks |
|--|-----------|-------------------------------------|
| <i>Escherichia coli</i> ATCC® 25922 | Inhibited | - |
| <i>Staphylococcus aureus</i> ATCC® 25923 | Inhibited | - |
| <i>Enterococcus faecalis</i> ATCC® 29212 | Good | Brown to black colonies (Esculin +) |
| <i>Streptococcus bovis</i> ATCC® 33317 | Good | Brown to black colonies (Esculin +) |
| <i>Enterococcus faecalis</i> ATCC® 51229 | Good | Brown to black colonies (Esculin +) |

REFERENCES

- CORRY, J.E.L., G.D.W. CURTIS & R.M. BAIRD. (2003) Handbook of Culture Media for Food Microbiology. Elsevier Sci. B.V. Amsterdam. The Netherlands.
- DOWNES, F.P. & K. ITO (2001) Compendium of Methods for the Microbiological Examination of Foods. 4th ed. APHA. Washington.
- GUINEA, J., SANCHO, J. & PARES, R. (1979) Análisis Microbiológico de Aguas. Ed. Omega. Barcelona.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- MOSSEL, D.A.A., P.G.M. BUKER, J. ELDERING (1978) Streptokokken der Lancefield Gruppe D in Lebensmitteln und Trinkwasser. Arch. F. Lebensmittelhyg. 29:121-127.
- PASCUAL ANDERSON. Mª.Rº. (1992) Microbiología Alimentaria. Díaz de Santos, S.A. Madrid.

STORAGE

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).



Th. Geyer GmbH & Co. KG
 Domierstr. 4 – 8
 D-71272 Renningen
 Tel.: +49 7159 1637-0
 Fax: +49 7159 1637-710
 renningen@thgeyer.de
 www.thgeyer.de

BW-Bank (Swift/BIC SOLADEST800)
 IBAN DE85800501010002036302
 Postbank Stuttgart (Swift/BIC PBNKDEFFXXX)
 IBAN DE32800100700000020708
 Deutsche Bank (Swift/BIC DEUTDESSXXX)
 IBAN DE06800700700125518100

St.-Nr. 70093/40018 / USt-IdNr. DE147510304
 Amtsgericht Stuttgart / HRA-Nr. 254140
 Persönlich haftende Gesellschafterin:
 Geyer Beteiligungsgesellschaft mbH
 Amtsgericht Stuttgart / HRB-Nr. 252035
 Geschäftsführer: Lutz-Alexander Geyer /
 Oliver-Alexander Geyer / André Meise / Ralf Streicher