

TECHNICAL DATA SHEET

Article no. : 8620.0500

LACTOSE BROTH

SPECIFICATION

Medium for pre-enrichment and detection of enterobacteria and coliforms in milk and water according to ISO standards.

FORMULA * IN G/L

Peptone of gelatin.....	5.0
Meat extract.....	3.0
Lactose.....	5.0

Final pH 6.9 ±0.2 at 25 °C

*Adjusted and /or supplemented as required to meet performance criteria

DIRECTIONS

Suspend 13 g of powder in 1 l of distilled water. Heat to boiling. Distribute into tubes with fitted Durham tubes and sterilize by autoclaving at 121 °C for 15 minutes.

DESCRIPTION

Lactose Broth is a classical medium for use in the presumptive testing for coliforms and for the enrichment of *Salmonella*. This formulation is per the standards recommended by APHA, AWWA, USP-NF and ISO.

It is commonly used with Durham fermentation tubes for the detection of gas formation. If a specific volume of sample is to be inoculated this must be taken into consideration when making up the medium as the concentration must not be altered on addition of the inoculum.

Although it is not Eijkman's original formulation, this broth provides excellent results in assays of gas production at 45 °C, which is a characteristic of *Escherichia coli*.

While preparing this medium it is important to avoid overheating and to distribute it into tubes before sterilization.

QUALITY CONTROL

Incubation Temperature: 35 ± 2 °C

Incubation Time: 24 ± 2 h

Inoculum: ≤100 CFU min. 50 CFU (productivity), according to ISO 11133:2014/Amd 1:2018.

Microorganism	Growth	Remarks
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Good	Gas (-). Durham tube
<i>Enterococcus faecalis</i> ATCC® 29212	Good	Gas (-). Durham tube
<i>Escherichia coli</i> ATCC® 8739	Good	Gas (+). Durham tube
<i>Escherichia coli</i> ATCC® 25922	Good	Gas (+). Durham tube
<i>Salmonella typhimurium</i> ATCC® 14028	Good	Gas (-). Durham tube
<i>Citrobacter freundii</i> ATCC® 43864	Good	Gas (+). Durham tube



Th. Geyer GmbH & Co. KG
Domierstr. 4 – 8
D-71272 Renningen
Tel.: +49 7159 1637-0
Fax: +49 7159 1637-710
renningen@thgeyer.de
www.thgeyer.de

BW-Bank (Swift/BIC SOLADEST800)
IBAN DE85800501010002036302
Postbank Stuttgart (Swift/BIC PBNKDEFFXXX)
IBAN DE32800100700000020708
Deutsche Bank (Swift/BIC DEUTDESSXXX)
IBAN DE06800700700125518100

St.-Nr. 70093/40018 / USt-IdNr. DE147510304
Amtsgericht Stuttgart / HRA-Nr. 254140
Persönlich haftende Gesellschafterin:
Geyer Beteiligungsgesellschaft mbH
Amtsgericht Stuttgart / HRB-Nr. 252035
Geschäftsführer: Lutz-Alexander Geyer /
Oliver-Alexander Geyer / André Meise / Ralf Streicher

REFERENCES

- APHA-AWWA-WPCF (1998) Standard methods for the examination of water and wastewater. 20th ed. APHA Washington.
- DOWNES, F.P. & K. ITO (2001) Compendium of Methods for the Microbiological Examination of Foods. 4th ed. APHA. Washington.
- FDA (Food and Drug Adminstrations) (1998) Bacteriological Analytical Manual 8th ed. Rev A. AOAC International. Gaithersburg. VA. USA.
- ISO 9308-2 Standard. (1990) Water Quality - Detection and enumeration of coliform organisms, thermotolerant coliform and presumptive E. coli - MPN technique.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 21150:2006 Standard. Cosmetics - Detection of Escherichia coli.
- US PHARMACOPOEIA (2005) <61> Microbial limit test. US Pharmacopeial Conv. Inc. Rockville. MD. USA.
- VANDERZANT & SPLITSTOESSER (1992) Compendium of Methods for the Microbiological Examination of Foods. 3rd ed. APHA. Washington.

STORAGE

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).



Th. Geyer GmbH & Co. KG
Domierstr. 4 – 8
D-71272 Renningen
Tel.: +49 7159 1637-0
Fax: +49 7159 1637-710
renningen@thgeyer.de
www.thgeyer.de

BW-Bank (Swift/BIC SOLADEST800)
IBAN DE85800501010002036302
Postbank Stuttgart (Swift/BIC PBNKDEFFXXX)
IBAN DE32800100700000020708
Deutsche Bank (Swift/BIC DEUTDESSXXX)
IBAN DE06800700700125518100

St.-Nr. 70093/40018 / USt-IdNr. DE147510304
Amtsgericht Stuttgart / HRA-Nr. 254140
Persönlich haftende Gesellschafterin:
Geyer Beteiligungsgesellschaft mbH
Amtsgericht Stuttgart / HRB-Nr. 252035
Geschäftsführer: Lutz-Alexander Geyer /
Oliver-Alexander Geyer / André Meise / Ralf Streicher